

Continuing Calibration Report for SequenceID : FE072825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055046.D

Aliphatic C28-C40	13569433.000	120.000	17.463	22.490	113078.608	120488.802	6.150
Aliphatic EPH	43256454.000	360.000	3.324	22.490	120156.817	125173.222	4.008
Aliphatic C9-C12	7098023.000	60.000	3.324	6.959	118300.383	123975.587	4.578
Aliphatic C12-C16	5092912.000	40.000	6.960	10.413	127322.800	132914.920	4.207
Aliphatic C16-C21	7661007.000	60.000	10.414	13.792	127683.450	131135.111	2.632
Aliphatic C21-C28	9835079.000	80.000	13.793	17.462	122938.488	124755.813	1.457