

Continuing Calibration Report for SequenceID : FE072825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055055.D

Aliphatic C28-C40	14024627.000	120.000	17.463	22.490	116871.892	120488.802	3.002
Aliphatic EPH	43865425.000	360.000	3.324	22.490	121848.403	125173.222	2.656
Aliphatic C9-C12	7116058.000	60.000	3.324	6.959	118600.967	123975.587	4.335
Aliphatic C12-C16	5109090.000	40.000	6.960	10.413	127727.250	132914.920	3.903
Aliphatic C16-C21	7715073.000	60.000	10.414	13.792	128584.550	131135.111	1.945
Aliphatic C21-C28	9900577.000	80.000	13.793	17.462	123757.213	124755.813	0.800