

Continuing Calibration Report for SequenceID : FE072925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055058.D

Aliphatic C9-C12	7325862.000	60.000	3.325	6.960	122097.700	123975.587	1.515
Aliphatic C12-C16	5242820.000	40.000	6.961	10.412	131070.500	132914.920	1.388
Aliphatic C16-C21	7929976.000	60.000	10.413	13.792	132166.267	131135.111	-0.786
Aliphatic C21-C28	10186540.000	80.000	13.793	17.463	127331.750	124755.813	-2.065
Aliphatic C28-C40	14319282.000	120.000	17.464	22.491	119327.350	120488.802	0.964
Aliphatic EPH	45004480.000	360.000	3.325	22.491	125012.444	125173.222	0.128