

Continuing Calibration Report for SequenceID : FE072925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055065.D

Aliphatic C9-C12	7271021.000	60.000	3.325	6.960	121183.683	123975.587	2.252
Aliphatic C12-C16	5205576.000	40.000	6.961	10.412	130139.400	132914.920	2.088
Aliphatic C16-C21	7820807.000	60.000	10.413	13.792	130346.783	131135.111	0.601
Aliphatic C21-C28	10078718.000	80.000	13.793	17.463	125983.975	124755.813	-0.984
Aliphatic C28-C40	14383470.000	120.000	17.464	22.491	119862.250	120488.802	0.520
Aliphatic EPH	44759592.000	360.000	3.325	22.491	124332.200	125173.222	0.672