

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP072125

Review By	Rahul	Review On	7/22/2025 8:59:18 AM
Supervise By	Jagrut	Supervise On	7/22/2025 9:10:39 AM
SubDirectory	BP072125	HP Acquire Method	BNA_P
		HP Processing Method	bp072125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840		
CCC	SP6836		
Internal Standard/PEM	S12674,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP025194.D	21 Jul 2025 13:02	CG/JU	Ok
2	SSTDCCC040	BP025195.D	21 Jul 2025 13:44	CG/JU	Not Ok
3	SSTDICC2.5	BP025196.D	21 Jul 2025 14:25	CG/JU	Ok
4	SSTDICC005	BP025197.D	21 Jul 2025 15:06	CG/JU	Ok,M
5	SSTDICC010	BP025198.D	21 Jul 2025 15:47	CG/JU	Ok
6	SSTDICC020	BP025199.D	21 Jul 2025 16:29	CG/JU	Ok
7	SSTDICCC040	BP025200.D	21 Jul 2025 17:10	CG/JU	Ok
8	SSTDICC050	BP025201.D	21 Jul 2025 17:51	CG/JU	Ok,M
9	SSTDICC060	BP025202.D	21 Jul 2025 18:32	CG/JU	Ok
10	SSTDICC080	BP025203.D	21 Jul 2025 19:14	CG/JU	Ok,M
11	SSTDICV040	BP025204.D	21 Jul 2025 19:55	CG/JU	Ok,M
12	PB168854BL	BP025205.D	21 Jul 2025 21:58	CG/JU	Ok

M : Manual Integration