

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112596.D **Time Analyzed:** 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.752	4.655	4.855	551.310	500.000	10.3
Aroclor-1016-2	4.770	4.674	4.874	541.610	500.000	8.3
Aroclor-1016-3	4.828	4.731	4.931	531.220	500.000	6.2
Aroclor-1016-4	4.947	4.851	5.051	540.490	500.000	8.1
Aroclor-1016-5	5.204	5.108	5.308	554.760	500.000	11.0
Aroclor-1260-1	6.241	6.145	6.345	546.820	500.000	9.4
Aroclor-1260-2	6.430	6.333	6.533	534.520	500.000	6.9
Aroclor-1260-3	6.796	6.700	6.900	565.580	500.000	13.1
Aroclor-1260-4	7.055	6.959	7.159	532.660	500.000	6.5
Aroclor-1260-5	7.298	7.203	7.403	552.810	500.000	10.6
Decachlorobiphenyl	8.690	8.593	8.793	51.320	50.000	2.6
Tetrachloro-m-xylene	3.667	3.569	3.769	54.600	50.000	9.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112596.D **Time Analyzed:** 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.735	4.638	4.838	501.770	500.000	0.4
Aroclor-1016-2	4.753	4.656	4.856	497.880	500.000	-0.4
Aroclor-1016-3	4.928	4.831	5.031	498.390	500.000	-0.3
Aroclor-1016-4	4.970	4.873	5.073	499.810	500.000	0.0
Aroclor-1016-5	5.182	5.085	5.285	506.230	500.000	1.2
Aroclor-1260-1	6.210	6.113	6.313	518.450	500.000	3.7
Aroclor-1260-2	6.399	6.302	6.502	511.070	500.000	2.2
Aroclor-1260-3	6.550	6.453	6.653	524.740	500.000	4.9
Aroclor-1260-4	7.019	6.923	7.123	511.890	500.000	2.4
Aroclor-1260-5	7.262	7.164	7.364	529.210	500.000	5.8
Decachlorobiphenyl	8.636	8.540	8.740	50.710	50.000	1.4
Tetrachloro-m-xylene	3.661	3.562	3.762	51.120	50.000	2.2