

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112610.D **Time Analyzed:** 15:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.755	4.655	4.855	534.540	500.000	6.9
Aroclor-1016-2	4.774	4.674	4.874	536.980	500.000	7.4
Aroclor-1016-3	4.831	4.731	4.931	525.870	500.000	5.2
Aroclor-1016-4	4.950	4.851	5.051	526.690	500.000	5.3
Aroclor-1016-5	5.207	5.108	5.308	575.840	500.000	15.2
Aroclor-1260-1	6.244	6.145	6.345	535.110	500.000	7.0
Aroclor-1260-2	6.434	6.333	6.533	530.010	500.000	6.0
Aroclor-1260-3	6.800	6.700	6.900	564.180	500.000	12.8
Aroclor-1260-4	7.059	6.959	7.159	539.520	500.000	7.9
Aroclor-1260-5	7.302	7.203	7.403	544.970	500.000	9.0
Decachlorobiphenyl	8.692	8.593	8.793	52.420	50.000	4.8
Tetrachloro-m-xylene	3.670	3.569	3.769	52.500	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112610.D **Time Analyzed:** 15:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.638	4.838	492.570	500.000	-1.5
Aroclor-1016-2	4.754	4.656	4.856	493.670	500.000	-1.3
Aroclor-1016-3	4.929	4.831	5.031	500.490	500.000	0.1
Aroclor-1016-4	4.972	4.873	5.073	498.480	500.000	-0.3
Aroclor-1016-5	5.184	5.085	5.285	517.310	500.000	3.5
Aroclor-1260-1	6.211	6.113	6.313	519.280	500.000	3.9
Aroclor-1260-2	6.400	6.302	6.502	518.200	500.000	3.6
Aroclor-1260-3	6.552	6.453	6.653	532.140	500.000	6.4
Aroclor-1260-4	7.021	6.923	7.123	525.920	500.000	5.2
Aroclor-1260-5	7.264	7.164	7.364	535.930	500.000	7.2
Decachlorobiphenyl	8.638	8.540	8.740	51.120	50.000	2.2
Tetrachloro-m-xylene	3.662	3.562	3.762	50.490	50.000	1.0