

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112629.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.655	4.855	523.150	500.000	4.6
Aroclor-1016-2	4.772	4.674	4.874	525.250	500.000	5.1
Aroclor-1016-3	4.828	4.731	4.931	515.170	500.000	3.0
Aroclor-1016-4	4.948	4.851	5.051	519.310	500.000	3.9
Aroclor-1016-5	5.205	5.108	5.308	539.090	500.000	7.8
Aroclor-1260-1	6.242	6.145	6.345	530.160	500.000	6.0
Aroclor-1260-2	6.432	6.333	6.533	527.480	500.000	5.5
Aroclor-1260-3	6.798	6.700	6.900	567.390	500.000	13.5
Aroclor-1260-4	7.057	6.959	7.159	537.760	500.000	7.6
Aroclor-1260-5	7.300	7.203	7.403	546.470	500.000	9.3
Decachlorobiphenyl	8.691	8.593	8.793	54.230	50.000	8.5
Tetrachloro-m-xylene	3.668	3.569	3.769	51.420	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112629.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	482.260	500.000	-3.5
Aroclor-1016-2	4.752	4.656	4.856	482.180	500.000	-3.6
Aroclor-1016-3	4.928	4.831	5.031	486.260	500.000	-2.7
Aroclor-1016-4	4.970	4.873	5.073	489.640	500.000	-2.1
Aroclor-1016-5	5.182	5.085	5.285	492.530	500.000	-1.5
Aroclor-1260-1	6.210	6.113	6.313	504.810	500.000	1.0
Aroclor-1260-2	6.398	6.302	6.502	505.920	500.000	1.2
Aroclor-1260-3	6.550	6.453	6.653	522.020	500.000	4.4
Aroclor-1260-4	7.020	6.923	7.123	516.970	500.000	3.4
Aroclor-1260-5	7.262	7.164	7.364	533.760	500.000	6.8
Decachlorobiphenyl	8.637	8.540	8.740	51.090	50.000	2.2
Tetrachloro-m-xylene	3.660	3.562	3.762	49.100	50.000	-1.8