

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112640.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.655	4.855	515.560	500.000	3.1
Aroclor-1016-2	4.773	4.674	4.874	510.730	500.000	2.1
Aroclor-1016-3	4.830	4.731	4.931	497.700	500.000	-0.5
Aroclor-1016-4	4.950	4.851	5.051	503.190	500.000	0.6
Aroclor-1016-5	5.206	5.108	5.308	545.690	500.000	9.1
Aroclor-1260-1	6.244	6.145	6.345	514.060	500.000	2.8
Aroclor-1260-2	6.434	6.333	6.533	513.070	500.000	2.6
Aroclor-1260-3	6.799	6.700	6.900	543.440	500.000	8.7
Aroclor-1260-4	7.059	6.959	7.159	512.500	500.000	2.5
Aroclor-1260-5	7.301	7.203	7.403	516.320	500.000	3.3
Decachlorobiphenyl	8.693	8.593	8.793	52.340	50.000	4.7
Tetrachloro-m-xylene	3.669	3.569	3.769	50.280	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2711
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112640.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.638	4.838	469.120	500.000	-6.2
Aroclor-1016-2	4.754	4.656	4.856	475.080	500.000	-5.0
Aroclor-1016-3	4.929	4.831	5.031	475.210	500.000	-5.0
Aroclor-1016-4	4.972	4.873	5.073	477.660	500.000	-4.5
Aroclor-1016-5	5.183	5.085	5.285	490.100	500.000	-2.0
Aroclor-1260-1	6.211	6.113	6.313	495.620	500.000	-0.9
Aroclor-1260-2	6.400	6.302	6.502	494.390	500.000	-1.1
Aroclor-1260-3	6.551	6.453	6.653	506.210	500.000	1.2
Aroclor-1260-4	7.021	6.923	7.123	506.310	500.000	1.3
Aroclor-1260-5	7.263	7.164	7.364	521.720	500.000	4.3
Decachlorobiphenyl	8.637	8.540	8.740	49.950	50.000	-0.1
Tetrachloro-m-xylene	3.661	3.562	3.762	48.110	50.000	-3.8