

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112412.D	23 Jul 2025 10:56		YP/AJ	Ok
2	I.BLK	I.BLK	PO112413.D	23 Jul 2025 11:14		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO112414.D	23 Jul 2025 11:32		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO112415.D	23 Jul 2025 11:50		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO112416.D	23 Jul 2025 12:08		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO112417.D	23 Jul 2025 12:27		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO112418.D	23 Jul 2025 12:45		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO112419.D	23 Jul 2025 13:03		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO112420.D	23 Jul 2025 13:22		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO112421.D	23 Jul 2025 13:40		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO112422.D	23 Jul 2025 13:59		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO112423.D	23 Jul 2025 14:17		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO112424.D	23 Jul 2025 14:36		YP/AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PO112425.D	23 Jul 2025 14:54		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO112426.D	23 Jul 2025 15:13		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO112427.D	23 Jul 2025 15:31		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO112428.D	23 Jul 2025 15:48		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO112429.D	23 Jul 2025 16:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO112430.D	23 Jul 2025 16:25		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO112431.D	23 Jul 2025 16:44		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO112432.D	23 Jul 2025 17:02		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO112433.D	23 Jul 2025 17:21		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO112434.D	23 Jul 2025 17:39		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO112435.D	23 Jul 2025 17:57		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO112436.D	23 Jul 2025 18:16		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO112437.D	23 Jul 2025 18:34		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO112438.D	23 Jul 2025 18:53		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO112439.D	23 Jul 2025 19:11		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO112440.D	23 Jul 2025 19:28		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO112441.D	23 Jul 2025 19:47		YP/AJ	Ok
31	PO072325ICV500	ICVPO072325	PO112442.D	23 Jul 2025 20:05		YP/AJ	Ok
32	AR1242ICV500	ICVPO072325AR1242	PO112443.D	23 Jul 2025 20:42		YP/AJ	Ok
33	AR1248ICV500	ICVPO072325AR1248	PO112444.D	23 Jul 2025 21:18		YP/AJ	Ok
34	AR1254ICV500	ICVPO072325AR1254	PO112445.D	23 Jul 2025 21:54		YP/AJ	Ok
35	AR1268ICV500	ICVPO072325AR1268	PO112446.D	23 Jul 2025 22:31		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO112447.D	23 Jul 2025 23:07		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO073025

Review By	yogesh	Review On	7/30/2025 11:18:03 AM
Supervise By	mohammad	Supervise On	8/1/2025 7:55:44 AM
SubDirectory	PO073025	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112595.D	30 Jul 2025 08:20		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112596.D	30 Jul 2025 08:37		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO112597.D	30 Jul 2025 08:56		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112598.D	30 Jul 2025 09:14		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO112599.D	30 Jul 2025 09:32		YP/AJ	Ok
6	I.BLK	I.BLK	PO112600.D	30 Jul 2025 09:56		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112601.D	30 Jul 2025 10:13		YP/AJ	Ok
8	Q2701-01	PIPE-1	PO112602.D	30 Jul 2025 10:35	AR1254 Hit	YP/AJ	Ok
9	PB169057BL	PB169057BL	PO112603.D	30 Jul 2025 12:18		YP/AJ	Ok
10	PB169057BS	PB169057BS	PO112604.D	30 Jul 2025 12:36		YP/AJ	Ok
11	Q2711-01	PCB15-33	PO112605.D	30 Jul 2025 12:55	AR1254 Hit	YP/AJ	Ok,M
12	Q2711-02	PCB15-31	PO112606.D	30 Jul 2025 13:13	AR1254 Hit	YP/AJ	Ok
13	Q2711-03	PCB15-28	PO112607.D	30 Jul 2025 13:32	AR1254 +1260 Hit,Carry over	YP/AJ	Not Ok
14	Q2711-04	PCB14-3	PO112608.D	30 Jul 2025 13:50	AR1254 Hit	YP/AJ	Ok
15	Q2711-05	PCB14-5	PO112609.D	30 Jul 2025 14:09	AR1254 Hit	YP/AJ	Ok
16	AR1660CCC500	AR1660CCC500	PO112610.D	30 Jul 2025 15:23		YP/AJ	Ok
17	AR1242CCC500	AR1242CCC500	PO112611.D	30 Jul 2025 16:19		YP/AJ	Ok
18	AR1248CCC500	AR1248CCC500	PO112612.D	30 Jul 2025 16:37		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO073025

Review By	yogesh	Review On	7/30/2025 11:18:03 AM
Supervise By	mohammad	Supervise On	8/1/2025 7:55:44 AM
SubDirectory	PO073025	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO112613.D	30 Jul 2025 16:56		YP/AJ	Ok
20	I.BLK	I.BLK	PO112614.D	30 Jul 2025 17:14		YP/AJ	Ok
21	PB169058BL	PB169058BL	PO112615.D	30 Jul 2025 17:33		YP/AJ	Ok
22	PB169058BS	PB169058BS	PO112616.D	30 Jul 2025 17:52		YP/AJ	Ok
23	Q2701-08DL	PIPE-8DL	PO112617.D	30 Jul 2025 18:10	AR1248 hit ,not required	YP/AJ	Not Ok
24	Q2690-01DL	723-ADL	PO112618.D	30 Jul 2025 18:27	AR1248 Hit	YP/AJ	Ok
25	Q2690-02DL	723-BDL	PO112619.D	30 Jul 2025 18:46	AR1248 Hit	YP/AJ	Ok
26	Q2701-03DL	PIPE-3DL	PO112620.D	30 Jul 2025 19:03	AR1248 +1254 HIT	YP/AJ	Ok,M
27	Q2722-01	LAW-25-0107	PO112621.D	30 Jul 2025 19:21	DCB high in both column	YP/AJ	ReRun
28	Q2722-03	LAW-25-0108	PO112622.D	30 Jul 2025 19:40	DCB high in both column	YP/AJ	ReRun
29	AR1660CCC500	AR1660CCC500	PO112623.D	30 Jul 2025 20:51		YP/AJ	Ok
30	AR1242CCC500	AR1242CCC500	PO112624.D	30 Jul 2025 21:47		YP/AJ	Ok
31	AR1248CCC500	AR1248CCC500	PO112625.D	30 Jul 2025 22:05		YP/AJ	Ok
32	AR1254CCC500	AR1254CCC500	PO112626.D	30 Jul 2025 22:24		YP/AJ	Ok
33	I.BLK	I.BLK	PO112627.D	30 Jul 2025 22:42		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO073125

Review By	yogesh	Review On	7/31/2025 10:22:06 AM
Supervise By	mohammad	Supervise On	8/1/2025 7:55:52 AM
SubDirectory	PO073125	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112628.D	31 Jul 2025 08:36		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112629.D	31 Jul 2025 08:54		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO112630.D	31 Jul 2025 09:14		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112631.D	31 Jul 2025 09:33		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO112632.D	31 Jul 2025 09:51		YP/AJ	Ok
6	I.BLK	I.BLK	PO112633.D	31 Jul 2025 10:09		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112634.D	31 Jul 2025 10:28		YP/AJ	Ok
8	Q2722-01RE	LAW-25-0107	PO112635.D	31 Jul 2025 10:46	DCB high in both column	YP/AJ	Confirms
9	Q2722-03	LAW-25-0108	PO112636.D	31 Jul 2025 11:05		YP/AJ	Ok,M
10	Q2701-08	PIPE-8	PO112637.D	31 Jul 2025 11:23	AR1248 Hit	YP/AJ	Ok
11	Q2690-01	723-A	PO112638.D	31 Jul 2025 12:00	AR1248 Hit, need 10x dilution	YP/AJ	Dilution
12	Q2711-03	PCB15-28	PO112639.D	31 Jul 2025 12:19	AR1260 Hit	YP/AJ	Ok,M
13	AR1660CCC500	AR1660CCC500	PO112640.D	31 Jul 2025 13:32		YP/AJ	Ok
14	AR1242CCC500	AR1242CCC500	PO112641.D	31 Jul 2025 13:50		YP/AJ	Ok
15	AR1248CCC500	AR1248CCC500	PO112642.D	31 Jul 2025 14:08		YP/AJ	Ok
16	AR1254CCC500	AR1254CCC500	PO112643.D	31 Jul 2025 14:27		YP/AJ	Ok
17	I.BLK	I.BLK	PO112644.D	31 Jul 2025 14:44		YP/AJ	Ok
18	PB169064BL	PB169064BL	PO112645.D	31 Jul 2025 15:02		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO073125

Review By	yogesh	Review On	7/31/2025 10:22:06 AM
Supervise By	mohammad	Supervise On	8/1/2025 7:55:52 AM
SubDirectory	PO073125	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB169064BS	PB169064BS	PO112646.D	31 Jul 2025 15:20		YP/AJ	Ok
20	Q2726-01	WBR-PBO	PO112647.D	31 Jul 2025 15:38		YP/AJ	Ok
21	Q2728-01	7211931-VNJ240	PO112648.D	31 Jul 2025 15:56		YP/AJ	Ok
22	Q2728-01MS	7211931-VNJ240MS	PO112649.D	31 Jul 2025 16:15		YP/AJ	Ok
23	Q2728-01MSD	7211931-VNJ240MSD	PO112650.D	31 Jul 2025 16:33		YP/AJ	Ok
24	AR1660CCC500	AR1660CCC500	PO112651.D	31 Jul 2025 17:47		YP/AJ	Ok
25	AR1242CCC500	AR1242CCC500	PO112652.D	31 Jul 2025 18:05		YP/AJ	Ok
26	AR1248CCC500	AR1248CCC500	PO112653.D	31 Jul 2025 18:23		YP/AJ	Ok
27	AR1254CCC500	AR1254CCC500	PO112654.D	31 Jul 2025 18:41		YP/AJ	Ok
28	I.BLK	I.BLK	PO112655.D	31 Jul 2025 19:00		YP/AJ	Ok

M : Manual Integration