

Surrogate Summary

SDG No.: Q2711

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	19.5	98		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
I.BLK-PO112600.D	PIBLK-PO112600.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	18.2	91		60	140
		Tetrachloro-m-xyl	2	20	17.7	89		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
PB169057BL	PB169057BL	Tetrachloro-m-xyl	1	20	20.2	101		32	144
		Decachlorobiphen	1	20	20.1	101		32	175
		Tetrachloro-m-xyl	2	20	18.9	95		32	144
		Decachlorobiphen	2	20	20.0	100		32	175
PB169057BS	PB169057BS	Tetrachloro-m-xyl	1	20	18.7	93		32	144
		Decachlorobiphen	1	20	19.3	96		32	175
		Tetrachloro-m-xyl	2	20	17.5	88		32	144
		Decachlorobiphen	2	20	19.3	96		32	175
Q2711-01	PCB15-33	Tetrachloro-m-xyl	1	20	19.7	99		32	144
		Decachlorobiphen	1	20	20.2	101		32	175
		Tetrachloro-m-xyl	2	20	18.9	95		32	144
		Decachlorobiphen	2	20	20.7	104		32	175
Q2711-02	PCB15-31	Tetrachloro-m-xyl	1	20	20.1	101		32	144
		Decachlorobiphen	1	20	21.2	106		32	175
		Tetrachloro-m-xyl	2	20	19.1	96		32	144
		Decachlorobiphen	2	20	21.1	106		32	175
Q2711-04	PCB14-3	Tetrachloro-m-xyl	1	20	20.4	102		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	19.5	98		32	144
		Decachlorobiphen	2	20	20.6	103		32	175
Q2711-05	PCB14-5	Tetrachloro-m-xyl	1	20	22.3	111		32	144
		Decachlorobiphen	1	20	20.3	102		32	175
		Tetrachloro-m-xyl	2	20	20.0	100		32	144
		Decachlorobiphen	2	20	21.0	105		32	175
I.BLK-PO112614.D	PIBLK-PO112614.D	Tetrachloro-m-xyl	1	20	18.5	93		60	140
		Decachlorobiphen	1	20	19.3	96		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	19.2	96		60	140
I.BLK-PO112633.D	PIBLK-PO112633.D	Tetrachloro-m-xyl	1	20	16.9	85		60	140
		Decachlorobiphen	1	20	18.9	94		60	140
		Tetrachloro-m-xyl	2	20	16.5	82		60	140
		Decachlorobiphen	2	20	18.5	92		60	140
Q2711-03	PCB15-28	Tetrachloro-m-xyl	1	20	19.7	98		32	144

Surrogate Summary

SDG No.: Q2711

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2711-03	PCB15-28	Decachlorobiphen	1	20	20.5	103		32	175
		Tetrachloro-m-xyl	2	20	18.3	92		32	144
		Decachlorobiphen	2	20	20.4	102		32	175
I.BLK-PO112644.D	PIBLK-PO112644.D	Tetrachloro-m-xyl	1	20	17.8	89		60	140
		Decachlorobiphen	1	20	19.1	95		60	140
		Tetrachloro-m-xyl	2	20	17.2	86		60	140
		Decachlorobiphen	2	20	18.5	92		60	140