

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GFEL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2712</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>07/29/2025 08:28</u>
Lab File ID:	<u>VL042773.D</u>	Init. Calib. Date(s):	<u>07/23/2025 07/23/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>08:58 13:00</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.528	0.450		-14.77	30
1,1-Dichloroethene	0.468	0.422		-9.83	30
cis-1,2-Dichloroethene	1.378	1.349		-2.1	30
1,1,1-Trichloroethane	2.158	2.076		-3.8	30
Trichloroethene	0.455	0.441		-3.08	30
Tetrachloroethene	0.393	0.370		-5.85	30
1-Bromo-4-Fluorobenzene	0.792	0.800		1.01	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.