

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072325

Review By	Semsettin Yesilyurt	Review On	7/24/2025 9:04:14 AM
Supervise By	Maresh Dadoda	Supervise On	7/24/2025 2:21:30 PM
SubDirectory	VL072325	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL072325AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2645
Initial Calibration Stds	AP2641,AP2642,AP2643
CCC	AP2641
Internal Standard/PEM	AP2645
ICV/I.BLK	AP2644
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042746.D	23 Jul 2025 08:00		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042747.D	23 Jul 2025 08:58		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042748.D	23 Jul 2025 09:34		SY/MD	Ok,M
4	VSTDICC001	VSTDICC001	VL042749.D	23 Jul 2025 10:08	Comp. #13,79,80,81 are on Linear Regression	SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042750.D	23 Jul 2025 10:42		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042751.D	23 Jul 2025 11:16	Not used	SY/MD	Not Ok
7	VSTDICC0.03	VSTDICC0.03	VL042752.D	23 Jul 2025 11:50		SY/MD	Ok,M
8	VSTDICC015	VSTDICC015	VL042753.D	23 Jul 2025 12:26	%D failed for Comp. #80 in 0.5 and 2 ppb	SY/MD	Ok,M
9	VSTDICC0.1	VSTDICC0.1	VL042754.D	23 Jul 2025 13:00	%D failed for Comp. #79 in 0.1 ppb	SY/MD	Ok,M
10	VSTDICV010	ICVVVL072325	VL042755.D	23 Jul 2025 14:33		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL072925

Review By	Semsettin Yesilyurt	Review On	7/30/2025 8:00:28 AM
Supervise By	Maresh Dadoda	Supervise On	7/30/2025 3:10:33 PM
SubDirectory	VL072925	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL072325AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2645		
Initial Calibration Stds	AP2641,AP2642,AP2643		
CCC	AP2641		
Internal Standard/PEM	AP2645		
ICV/I.BLK	AP2644		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042772.D	29 Jul 2025 07:37		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL042773.D	29 Jul 2025 08:28		SY/MD	Ok,M
3	VL0729ABL01	VL0729ABL01	VL042774.D	29 Jul 2025 09:16		SY/MD	Ok
4	VL0729ABS01	VL0729ABS01	VL042775.D	29 Jul 2025 10:03		SY/MD	Ok,M
5	Q2712-01	SV1	VL042776.D	29 Jul 2025 10:50	Need 40X	SY/MD	Dilution
6	Q2712-01DL	SV1DL	VL042777.D	29 Jul 2025 11:26		SY/MD	Ok,M
7	Q2712-02	SV2	VL042778.D	29 Jul 2025 12:02		SY/MD	Ok,M
8	Q2712-02DL	SV2DL	VL042779.D	29 Jul 2025 12:37	Not Required	SY/MD	Not Ok
9	Q2712-03	SV3	VL042780.D	29 Jul 2025 13:13		SY/MD	Ok,M
10	Q2712-03DL	SV3DL	VL042781.D	29 Jul 2025 13:49	Not Required	SY/MD	Not Ok
11	Q2712-04	IA2	VL042782.D	29 Jul 2025 14:24		SY/MD	Ok,M
12	Q2712-04DUP	IA2DUP	VL042783.D	29 Jul 2025 14:59		SY/MD	Ok,M
13	Q2712-05	IA3	VL042784.D	29 Jul 2025 15:35		SY/MD	Ok,M
14	Q2712-06	OA1	VL042785.D	29 Jul 2025 16:10		SY/MD	Ok,M
15	Q2712-07	IA1	VL042786.D	29 Jul 2025 16:48		SY/MD	Ok,M

M : Manual Integration