

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136651

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.092	0.099	93	85-115	07/30/2025
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	07/30/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	07/30/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136682

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.0	7	100	90-110	07/29/2025
Sample ID:	CCV1						
Field pH		pH	7.0	7.0	100	90-110	07/29/2025
Sample ID:	CCV2						
Field pH		pH	7.0	7.0	100	90-110	07/29/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136682

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------