

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136651

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0015	0.0025	J	0.0012	0.005	07/30/2025
Sample ID: CCB1 Cyanide	mg/L	0.0014	0.0025	J	0.0012	0.005	07/30/2025
Sample ID: CCB2 Cyanide	mg/L	0.0016	0.0025	J	0.0012	0.005	07/30/2025

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136651

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------