

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q2724

Project: Parsippany Wastewater Quarterly 2025

Run No.: LB136651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169055BS							
Cyanide	mg/L	0.1	0.093		93	1	85-115	07/30/2025