



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: METEM A GE POWER Business

SDG No.: Q2724

Contract: METE01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.67	6.0	94	65 - 135	P	07/31/2025	14:10	LB136666
	Chromium	10.1	10.0	101	65 - 135	P	07/31/2025	14:10	LB136666
	Copper	23.0	20.0	115	65 - 135	P	07/31/2025	14:10	LB136666
	Lead	11.5	12.0	96	65 - 135	P	07/31/2025	14:10	LB136666
	Nickel	39.6	40.0	99	65 - 135	P	07/31/2025	14:10	LB136666
	Silver	10.9	10.0	109	65 - 135	P	07/31/2025	14:10	LB136666
	Zinc	48.8	40.0	122	65 - 135	P	07/31/2025	14:10	LB136666