



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2724

Contract: METE01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 14:06            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 14:06            | LB136666      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2724

Contract: METE01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 15:11            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 15:11            | LB136666      |
| CCB02     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 16:02            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 16:02            | LB136666      |
| CCB03     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 16:53            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 16:53            | LB136666      |
| CCB04     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 17:24            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 17:24            | LB136666      |
| CCB05     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 18:25            | LB136666      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 18:25            | LB136666      |
| CCB06     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 19:22            | LB136666      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 19:22            | LB136666      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 19:22            | LB136666      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 19:22            | LB136666      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** METEM A GE POWER Business

**SDG No.:** Q2724

**Contract:** METE01

**Lab Code:** ACE

| Sample ID    | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|--------------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| <b>CCB06</b> | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 19:22            | LB136666      |
|              | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 19:22            | LB136666      |
|              | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 19:22            | LB136666      |
| <b>CCB07</b> | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/31/2025       | 20:04            | LB136666      |
|              | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/31/2025       | 20:04            | LB136666      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** METEM A GE POWER Business

**SDG No.:** Q2724

**Contract:** METE01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|