

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business

SDG No.: Q2724

Contract: METE01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	0.36	1.0	36	-5	7	07/31/2025	14:33	LB136666
	Chromium	54.9	52.0	106	42	62	07/31/2025	14:33	LB136666
	Copper	-2.66	2.0	133	-18	22	07/31/2025	14:33	LB136666
	Lead	-5.42			-12	12	07/31/2025	14:33	LB136666
	Nickel	3.34	2.0	167	-38	42	07/31/2025	14:33	LB136666
	Silver	2.78			-10	10	07/31/2025	14:33	LB136666
	Zinc	1.26			-40	40	07/31/2025	14:33	LB136666
ICSAB01	Cadmium	905	972	93	826	1120	07/31/2025	14:40	LB136666
	Chromium	522	542	96	460	624	07/31/2025	14:40	LB136666
	Copper	455	511	89	434	588	07/31/2025	14:40	LB136666
	Lead	40.2	49.0	82	37	61	07/31/2025	14:40	LB136666
	Nickel	935	954	98	810	1100	07/31/2025	14:40	LB136666
	Silver	207	201	103	170	232	07/31/2025	14:40	LB136666
	Zinc	996	952	105	809	1095	07/31/2025	14:40	LB136666
ICSA	Cadmium	-0.080	1.0	8	-5	7	07/31/2025	14:51	LB136666
	Chromium	40.9	52.0	79	42	62	07/31/2025	14:51	LB136666
	Copper	-3.18	2.0	159	-18	22	07/31/2025	14:51	LB136666
	Lead	4.57			-12	12	07/31/2025	14:51	LB136666
	Nickel	3.13	2.0	156	-38	42	07/31/2025	14:51	LB136666
	Silver	13.8			-10	10	07/31/2025	14:51	LB136666
	Zinc	-21.5			-40	40	07/31/2025	14:51	LB136666
ICSAB	Cadmium	894	972	92	826	1120	07/31/2025	14:56	LB136666
	Chromium	524	542	97	460	624	07/31/2025	14:56	LB136666
	Copper	490	511	96	434	588	07/31/2025	14:56	LB136666
	Lead	33.8	49.0	69	37	61	07/31/2025	14:56	LB136666
	Nickel	935	954	98	810	1100	07/31/2025	14:56	LB136666
	Silver	206	201	102	170	232	07/31/2025	14:56	LB136666
	Zinc	976	952	102	809	1095	07/31/2025	14:56	LB136666