

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2725

Project: Monthly 2025

Run No.: LB136657

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136657BS							
BOD5	mg/L	198	192		97	1	84.6-115.4	07/30/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2725

Project: Monthly 2025

Run No.: LB136660

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136660BS							
TPH	mg/L	20.0	17.0		85	1	78-114	07/31/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2725

Project: Monthly 2025

Run No.: LB136660

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136660BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	07/31/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2725

Project: Monthly 2025

Run No.: LB136662

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136662BS							
TSS	mg/L	550	533		97	1	90-110	07/31/2025