



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q2725

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.045	0.05	89	60 - 140	CV	07/31/2025	13:45	LB136665
CRI01	Cadmium	6.14	6.0	102	65 - 135	P	08/04/2025	11:57	LB136697
	Copper	22.9	20.0	115	65 - 135	P	08/04/2025	11:57	LB136697
	Lead	11.2	12.0	93	65 - 135	P	08/04/2025	11:57	LB136697
	Nickel	41.8	40.0	105	65 - 135	P	08/04/2025	11:57	LB136697
	Zinc	44.8	40.0	112	65 - 135	P	08/04/2025	11:57	LB136697