

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2725</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2725-02</u>	client id: <u>CompMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2725-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	98.1	D	15.0	UD	100	97		P
Copper	ug/L	75 - 125	347	D	197	D	150	100		P
Lead	ug/L	75 - 125	521	D	42.8	D	500	96		P
Mercury	ug/L	75 - 125	2.67		0.26		4.0	60	N	CV
Nickel	ug/L	75 - 125	274	D	25.0	JD	250	100		P
Zinc	ug/L	75 - 125	824	D	731	D	100	93		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>Q2725</u>
contract: <u>ARAM01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2725-02</u>	client id: <u>CompMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2725-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	94.6	D	15.0	UD	100	93		P
Copper	ug/L	75 - 125	334	D	197	D	150	91		P
Lead	ug/L	75 - 125	502	D	42.8	D	500	92		P
Mercury	ug/L	75 - 125	2.32		0.26		4.0	52	N	CV
Nickel	ug/L	75 - 125	264	D	25.0	JD	250	96		P
Zinc	ug/L	75 - 125	800	D	731	D	100	68		P