

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG No.: Q2727
 Instrument ID: MSVOA_L Calibration Date/Time: 07/31/2025 08:37
 Lab File ID: VL042788.D Init. Calib. Date(s): 07/23/2025 07/23/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 08:58 13:00
 GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Ethanol	10.000	10.15	-1.499996	20
1,2,4-Trichlorobenzene	10.000	10.99	-9.870005	20
Naphthalene	10.000	11.01	-10.14	20