

Surrogate Summary

SDG No.: Q2727

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2727-01	AE073-1100	1-Bromo-4-Fluorobenzene	10	10.7	107		65	135
Q2727-01DL	AE073-1100DL	1-Bromo-4-Fluorobenzene	10	10.6	106		65	135
Q2727-01DL2	AE073-1100DL2	1-Bromo-4-Fluorobenzene	10	9.96	100		65	135
VL0731ABL01	VL0731ABL01	1-Bromo-4-Fluorobenzene	10	9.82	98		65	135
VL0731ABS01	VL0731ABS01	1-Bromo-4-Fluorobenzene	10	10.3	103		65	135
VL0731ABSD01	VL0731ABSD01	1-Bromo-4-Fluorobenzene	10	10.4	104		65	135
VL0813ABL01	VL0813ABL01	1-Bromo-4-Fluorobenzene	10	10.0	100		65	135
VL0813ABS01	VL0813ABS01	1-Bromo-4-Fluorobenzene	10	10.1	101		65	135