

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q2727
 Lab File ID: VL042842.D Date Analyzed: 08/13/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:30
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	133104	2.79	413632	3.97	370372	8.89
UPPER LIMIT	186346	3.12	579085	4.30	518521	9.22
LOWER LIMIT	79862.4	2.46	248179	3.64	222223	8.56
EPA SAMPLE NO.						
AE073-1100DL2	118347	2.79	364683	3.97	318426	8.89
VL0813ABL01	119946	2.79	373094	3.97	324853	8.89
VL0813ABS01	116881	2.79	368200	3.97	322291	8.90

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.