



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Order ID # Q2729

Test Name: Metals Group 10

A. Number of Samples and Date of Receipt:

2 Water samples were received on 07/30/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group 10, TSS and VOC-BTEX. This data package contains results for Metals Group 10.

C. Analytical Techniques:

The analysis and digestion of Metals Group 10 was based on method 200.7.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all Compounds.

The Duplicate analysis met criteria for all Compounds.

The Matrix Spike analysis met criteria for all Compounds.

The Matrix Spike Duplicate analysis met criteria for all Compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The temperature of the samples at the time of receipt was 22.6°C.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____