

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1040	1000	104	95 - 105	P	07/31/2025	13:51	LB136666
	Iron	4130	4000	103	95 - 105	P	07/31/2025	13:51	LB136666
	Lead	3830	4000	96	95 - 105	P	07/31/2025	13:51	LB136666

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.3	20.0	107	80 - 120	P	07/31/2025	14:00	LB136666
	Iron	103	100	103	80 - 120	P	07/31/2025	14:00	LB136666
	Lead	10.2	12.0	85	80 - 120	P	07/31/2025	14:00	LB136666

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1220	1250	97	90 - 110	P	07/31/2025	15:06	LB136666
	Iron	4880	5000	98	90 - 110	P	07/31/2025	15:06	LB136666
	Lead	4760	5000	95	90 - 110	P	07/31/2025	15:06	LB136666
CCV02	Copper	1240	1250	99	90 - 110	P	07/31/2025	15:58	LB136666
	Iron	4960	5000	99	90 - 110	P	07/31/2025	15:58	LB136666
	Lead	4850	5000	97	90 - 110	P	07/31/2025	15:58	LB136666
CCV03	Copper	1230	1250	98	90 - 110	P	07/31/2025	16:49	LB136666
	Iron	4970	5000	99	90 - 110	P	07/31/2025	16:49	LB136666
	Lead	4830	5000	97	90 - 110	P	07/31/2025	16:49	LB136666
CCV04	Copper	1250	1250	100	90 - 110	P	07/31/2025	17:20	LB136666
	Iron	5010	5000	100	90 - 110	P	07/31/2025	17:20	LB136666
	Lead	4870	5000	98	90 - 110	P	07/31/2025	17:20	LB136666
CCV05	Copper	1230	1250	98	90 - 110	P	07/31/2025	18:20	LB136666
	Iron	5040	5000	101	90 - 110	P	07/31/2025	18:20	LB136666
	Lead	4830	5000	97	90 - 110	P	07/31/2025	18:20	LB136666
CCV06	Copper	1230	1250	98	90 - 110	P	07/31/2025	19:18	LB136666
	Iron	5010	5000	100	90 - 110	P	07/31/2025	19:18	LB136666
	Lead	4820	5000	96	90 - 110	P	07/31/2025	19:18	LB136666
CCV07	Copper	1240	1250	100	90 - 110	P	07/31/2025	20:00	LB136666
	Iron	5150	5000	103	90 - 110	P	07/31/2025	20:00	LB136666
	Lead	4880	5000	98	90 - 110	P	07/31/2025	20:00	LB136666