



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	07/31/2025	14:06	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	14:06	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	14:06	LB136666

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	07/31/2025	15:11	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	15:11	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	15:11	LB136666
CCB02	Copper	4.60	+/-10	U	20.0	P	07/31/2025	16:02	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	16:02	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	16:02	LB136666
CCB03	Copper	4.60	+/-10	U	20.0	P	07/31/2025	16:53	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	16:53	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	16:53	LB136666
CCB04	Copper	4.60	+/-10	U	20.0	P	07/31/2025	17:24	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	17:24	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	17:24	LB136666
CCB05	Copper	4.60	+/-10	U	20.0	P	07/31/2025	18:25	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	18:25	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	18:25	LB136666
CCB06	Copper	4.60	+/-10	U	20.0	P	07/31/2025	19:22	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	19:22	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	19:22	LB136666
CCB07	Copper	4.60	+/-10	U	20.0	P	07/31/2025	20:04	LB136666
	Iron	23.4	+/-50	U	100	P	07/31/2025	20:04	LB136666
	Lead	2.30	+/-6	U	12.0	P	07/31/2025	20:04	LB136666