

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-2.66	2.0	133	-18	22	07/31/2025	14:33	LB136666
	Iron	101000	101000	100	85600	116500	07/31/2025	14:33	LB136666
	Lead	-5.42			-12	12	07/31/2025	14:33	LB136666
ICSAB01	Copper	455	511	89	434	588	07/31/2025	14:40	LB136666
	Iron	100000	99300	101	84400	114500	07/31/2025	14:40	LB136666
	Lead	40.2	49.0	82	37	61	07/31/2025	14:40	LB136666
ICSA	Copper	-3.18	2.0	159	-18	22	07/31/2025	14:51	LB136666
	Iron	98300	101000	97	85600	116500	07/31/2025	14:51	LB136666
	Lead	4.57			-12	12	07/31/2025	14:51	LB136666
ICSAB	Copper	490	511	96	434	588	07/31/2025	14:56	LB136666
	Iron	94700	99300	95	84400	114500	07/31/2025	14:56	LB136666
	Lead	33.8	49.0	69	37	61	07/31/2025	14:56	LB136666