

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2729

Contract: TULL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169059BS							
Copper	ug/L	150	155		103	85 - 115	P
Iron	ug/L	1500	1520		101	85 - 115	P
Lead	ug/L	500	465		93	85 - 115	P