

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

SDG NO.: Q2729

Lab File ID: _____

BFB Injection Date: _____

Instrument ID: _____

BFB Injection Time: _____

GC Column: _____ ID: _____ (mm)

Heated Purge: Y/N _____

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50		
75		
95	Base Peak, 100% relative abundance	
96	5.0 - 9.0% of mass 95	
173	Less than 2.0% of mass 174	() 1
174		
175		() 1
176		() 1
177	5.0 - 9.0% of mass 176	() 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED