

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: Q2730

Project: Transfer Station-SPDES

RunNo.: LB136664

| Analyte | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|--|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: ICV1 Ammonia as N | mg/L | 1 | 1 | 100 | 90-110 | 07/31/2025 |
| Sample ID: CCV1 Ammonia as N | mg/L | 1 | 1 | 100 | 90-110 | 07/31/2025 |
| Sample ID: CCV2 Ammonia as N | mg/L | 1 | 1 | 100 | 90-110 | 07/31/2025 |
| Sample ID: CCV3 Ammonia as N | mg/L | 0.98 | 1 | 98 | 90-110 | 07/31/2025 |

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: Q2730

Project: Transfer Station-SPDES

RunNo.: LB136664

| Analyte | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|---------|-------|--------|------------|---------------|---------------------------|------------------|
|---------|-------|--------|------------|---------------|---------------------------|------------------|