

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q2730
Project:	Transfer Station-SPDES	Sample ID:	Q2698-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	28.2		28.4		1	0.71		07/31/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q2730
Project: Transfer Station-SPDES	Sample ID: Q2702-01
Client ID: MH-7-25-25MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	167		167		1	0.06		07/31/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q2730
Project:	Transfer Station-SPDES	Sample ID:	Q2725-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	744		750		1	0.8		07/31/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q2730
Project:	Transfer Station-SPDES	Sample ID:	Q2730-01
Client ID:	001 willets Pt Blvd(july)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	205		196		1	4.79		07/30/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q2730
Project: Transfer Station-SPDES	Sample ID: Q2730-02
Client ID: 002 35th Ave(july)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.40	OR	6.50	OR	1	2		07/31/2025
Ammonia as N	mg/L	+/-20	5.90	D	6.30	D	5	7		07/31/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q2730
Project: Transfer Station-SPDES	Sample ID: Q2730-02
Client ID: 002 35th Ave(july)MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	7.60	OR	7.60	OR	1	0		07/31/2025