

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	07/29/25
Project:	Transfer Station-SPDES	Date Received:	07/30/25
Client Sample ID:	001 willets Pt Blvd(july)	SDG No.:	Q2730
Lab Sample ID:	Q2730-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua. DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	51.8	1	1.89	10.0	ug/L	07/30/25 14:05	07/31/25 16:37	EPA 200.7	
7439-89-6	Iron	10400	1	18.0	50.0	ug/L	07/30/25 14:05	07/31/25 16:37	EPA 200.7	
7439-92-1	Lead	42.7	1	1.21	6.00	ug/L	07/30/25 14:05	07/31/25 16:37	EPA 200.7	

Color Before: Colorless Clarity Before: Clear Texture:

Color After: Colorless Clarity After: Clear Artifacts:

Comments: Metals Group 10

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence

of interference.

OR = Over Range

N =Spiked sample recovery not within control limits