

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2730</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2724-01</u>	client id: <u>WASTEWATER Q3MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2724-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	440		280		150	107		P
Iron	ug/L	75 - 125	1530		50.0	U	1500	102		P
Lead	ug/L	75 - 125	430		6.00	U	500	86		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2730</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2724-01</u>	client id: <u>WASTEWATER Q3MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2724-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	427		280		150	98		P
Iron	ug/L	75 - 125	1440		50.0	U	1500	96		P
Lead	ug/L	75 - 125	399		6.00	U	500	80		P