

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2730
 Lab File ID: VN087437.D Date Analyzed: 07/30/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:19
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001 willets Pt Blvd(july)	0	0.00				
001 willets Pt Blvd(july)DL	0	0.00				
002 35th Ave(july)	0	0.00				
002 35th Ave(july)DL	0	0.00				
VN0730WBL01	0	0.00				
VN0730WBS01	0	0.00				
VN0730WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.