



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ENTACT

Project Name: 540 Degraw St, Brooklyn, NY - E9309

Project # N/A

Order ID # Q2732

Test Name: TCLP Mercury, TCLP Metals Group2

A. Number of Samples and Date of Receipt:

1 Solid samples were received on 08/12/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, Corrosivity, Ignitability, Mercury, Metals Group1, Metals ICP-Group1, Oil and Grease, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TCLP Metals Group2, TS and TVS. This data package contains results for TCLP Mercury, TCLP Metals Group2.

C. Analytical Techniques:

The analysis of TCLP Metals Group2 was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate (TP-10DUP) analysis met criteria for all compounds except for Barium due to sample matrix interference.

The Matrix Spike (TP-10MS) analysis met criteria for all compounds except for Barium due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (TP-10MSD) analysis met criteria for all compounds except for Barium due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

As per special requirement for this project form-1 and Hit summary are reported in mg/L.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____