

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV14	Mercury	3.96	4.0	99	90 - 110	CV	08/14/2025	10:58	LB136819

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV41	Mercury	5.13	5.0	102	90 - 110	CV	08/14/2025	11:03	LB136819
CCV42	Mercury	5.09	5.0	102	90 - 110	CV	08/14/2025	11:34	LB136819
CCV43	Mercury	5.03	5.0	100	90 - 110	CV	08/14/2025	12:01	LB136819
CCV44	Mercury	4.98	5.0	100	90 - 110	CV	08/14/2025	12:13	LB136819

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3980	4000	100	90 - 110	P	08/14/2025	12:05	LB136837
	Barium	8040	8000	100	90 - 110	P	08/14/2025	12:05	LB136837
	Cadmium	1980	2000	99	90 - 110	P	08/14/2025	12:05	LB136837
	Chromium	820	800	102	90 - 110	P	08/14/2025	12:05	LB136837
	Copper	1030	1000	102	90 - 110	P	08/14/2025	12:05	LB136837
	Lead	3890	4000	97	90 - 110	P	08/14/2025	12:05	LB136837
	Nickel	2010	2000	100	90 - 110	P	08/14/2025	12:05	LB136837
	Selenium	3970	4000	99	90 - 110	P	08/14/2025	12:05	LB136837
	Silver	1070	1000	107	90 - 110	P	08/14/2025	12:05	LB136837
	Zinc	2000	2000	100	90 - 110	P	08/14/2025	12:05	LB136837

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.4	20.0	107	80 - 120	P	08/14/2025	12:09	LB136837
	Barium	97.9	100	98	80 - 120	P	08/14/2025	12:09	LB136837
	Cadmium	5.85	6.0	97	80 - 120	P	08/14/2025	12:09	LB136837
	Chromium	10.7	10.0	107	80 - 120	P	08/14/2025	12:09	LB136837
	Copper	21.6	20.0	108	80 - 120	P	08/14/2025	12:09	LB136837
	Lead	12.2	12.0	101	80 - 120	P	08/14/2025	12:09	LB136837
	Nickel	40.2	40.0	100	80 - 120	P	08/14/2025	12:09	LB136837
	Selenium	19.6	20.0	98	80 - 120	P	08/14/2025	12:09	LB136837
	Silver	10.5	10.0	105	80 - 120	P	08/14/2025	12:09	LB136837
	Zinc	42.9	40.0	107	80 - 120	P	08/14/2025	12:09	LB136837

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5070	5000	102	90 - 110	P	08/14/2025	12:50	LB136837
	Barium	9910	10000	99	90 - 110	P	08/14/2025	12:50	LB136837
	Cadmium	2490	2500	100	90 - 110	P	08/14/2025	12:50	LB136837
	Chromium	1020	1000	102	90 - 110	P	08/14/2025	12:50	LB136837
	Copper	1270	1250	101	90 - 110	P	08/14/2025	12:50	LB136837
	Lead	4960	5000	99	90 - 110	P	08/14/2025	12:50	LB136837
	Nickel	2490	2500	100	90 - 110	P	08/14/2025	12:50	LB136837
	Selenium	5140	5000	103	90 - 110	P	08/14/2025	12:50	LB136837
	Silver	1270	1250	102	90 - 110	P	08/14/2025	12:50	LB136837
	Zinc	2550	2500	102	90 - 110	P	08/14/2025	12:50	LB136837
CCV02	Arsenic	5090	5000	102	90 - 110	P	08/14/2025	13:42	LB136837
	Barium	9920	10000	99	90 - 110	P	08/14/2025	13:42	LB136837
	Cadmium	2460	2500	98	90 - 110	P	08/14/2025	13:42	LB136837
	Chromium	1020	1000	102	90 - 110	P	08/14/2025	13:42	LB136837
	Copper	1260	1250	101	90 - 110	P	08/14/2025	13:42	LB136837
	Lead	4890	5000	98	90 - 110	P	08/14/2025	13:42	LB136837
	Nickel	2480	2500	99	90 - 110	P	08/14/2025	13:42	LB136837
	Selenium	5140	5000	103	90 - 110	P	08/14/2025	13:42	LB136837
	Silver	1260	1250	101	90 - 110	P	08/14/2025	13:42	LB136837
	Zinc	2580	2500	103	90 - 110	P	08/14/2025	13:42	LB136837
CCV03	Arsenic	4980	5000	100	90 - 110	P	08/14/2025	14:41	LB136837
	Barium	9880	10000	99	90 - 110	P	08/14/2025	14:41	LB136837
	Cadmium	2400	2500	96	90 - 110	P	08/14/2025	14:41	LB136837
	Chromium	984	1000	98	90 - 110	P	08/14/2025	14:41	LB136837
	Copper	1240	1250	99	90 - 110	P	08/14/2025	14:41	LB136837
	Lead	4790	5000	96	90 - 110	P	08/14/2025	14:41	LB136837
	Nickel	2420	2500	97	90 - 110	P	08/14/2025	14:41	LB136837
	Selenium	5000	5000	100	90 - 110	P	08/14/2025	14:41	LB136837
	Silver	1240	1250	99	90 - 110	P	08/14/2025	14:41	LB136837
	Zinc	2520	2500	101	90 - 110	P	08/14/2025	14:41	LB136837
CCV04	Arsenic	5090	5000	102	90 - 110	P	08/14/2025	15:58	LB136837
	Barium	9790	10000	98	90 - 110	P	08/14/2025	15:58	LB136837
	Cadmium	2410	2500	96	90 - 110	P	08/14/2025	15:58	LB136837
	Chromium	1000	1000	100	90 - 110	P	08/14/2025	15:58	LB136837

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1260	1250	100	90 - 110	P	08/14/2025	15:58	LB136837
	Lead	4810	5000	96	90 - 110	P	08/14/2025	15:58	LB136837
	Nickel	2440	2500	98	90 - 110	P	08/14/2025	15:58	LB136837
	Selenium	5210	5000	104	90 - 110	P	08/14/2025	15:58	LB136837
	Silver	1260	1250	101	90 - 110	P	08/14/2025	15:58	LB136837
	Zinc	2580	2500	103	90 - 110	P	08/14/2025	15:58	LB136837
CCV05	Arsenic	5110	5000	102	90 - 110	P	08/14/2025	16:50	LB136837
	Barium	9710	10000	97	90 - 110	P	08/14/2025	16:50	LB136837
	Cadmium	2390	2500	96	90 - 110	P	08/14/2025	16:50	LB136837
	Chromium	985	1000	98	90 - 110	P	08/14/2025	16:50	LB136837
	Copper	1260	1250	100	90 - 110	P	08/14/2025	16:50	LB136837
	Lead	4770	5000	96	90 - 110	P	08/14/2025	16:50	LB136837
	Nickel	2430	2500	97	90 - 110	P	08/14/2025	16:50	LB136837
	Selenium	5230	5000	104	90 - 110	P	08/14/2025	16:50	LB136837
	Silver	1250	1250	100	90 - 110	P	08/14/2025	16:50	LB136837
	Zinc	2590	2500	104	90 - 110	P	08/14/2025	16:50	LB136837
CCV06	Arsenic	5150	5000	103	90 - 110	P	08/14/2025	17:54	LB136837
	Barium	9970	10000	100	90 - 110	P	08/14/2025	17:54	LB136837
	Cadmium	2380	2500	95	90 - 110	P	08/14/2025	17:54	LB136837
	Chromium	984	1000	98	90 - 110	P	08/14/2025	17:54	LB136837
	Copper	1260	1250	100	90 - 110	P	08/14/2025	17:54	LB136837
	Lead	4750	5000	95	90 - 110	P	08/14/2025	17:54	LB136837
	Nickel	2430	2500	97	90 - 110	P	08/14/2025	17:54	LB136837
	Selenium	5260	5000	105	90 - 110	P	08/14/2025	17:54	LB136837
	Silver	1250	1250	100	90 - 110	P	08/14/2025	17:54	LB136837
	Zinc	2570	2500	103	90 - 110	P	08/14/2025	17:54	LB136837