



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB14	Mercury	0.076	+/-0.2	U	0.20	CV	08/14/2025	11:00	LB136819

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB41	Mercury	0.076	+/-0.2	U	0.20	CV	08/14/2025	11:05	LB136819
CCB42	Mercury	0.076	+/-0.2	U	0.20	CV	08/14/2025	11:36	LB136819
CCB43	Mercury	0.076	+/-0.2	U	0.20	CV	08/14/2025	12:04	LB136819
CCB44	Mercury	0.076	+/-0.2	U	0.20	CV	08/14/2025	12:15	LB136819

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	12:13	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	12:13	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	12:13	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	12:13	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	12:13	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	12:13	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	12:13	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	12:13	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	12:13	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	12:13	LB136837



- 3a -

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	12:54	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	12:54	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	12:54	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	12:54	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	12:54	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	12:54	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	12:54	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	12:54	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	12:54	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	12:54	LB136837
CCB02	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	13:46	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	13:46	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	13:46	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	13:46	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	13:46	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	13:46	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	13:46	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	13:46	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	13:46	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	13:46	LB136837
CCB03	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	14:45	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	14:45	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	14:45	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	14:45	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	14:45	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	14:45	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	14:45	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	14:45	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	14:45	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	14:45	LB136837
CCB04	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	16:02	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	16:02	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	16:02	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	16:02	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	16:02	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	16:02	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	16:02	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	16:02	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	16:02	LB136837

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2732

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	16:02	LB136837
CCB05	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	17:00	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	17:00	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	17:00	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	17:00	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	17:00	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	17:00	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	17:00	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	17:00	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	17:00	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	17:00	LB136837
CCB06	Arsenic	5.12	+/-10	U	20.0	P	08/14/2025	18:05	LB136837
	Barium	14.6	+/-50	U	100	P	08/14/2025	18:05	LB136837
	Cadmium	0.50	+/-3	U	6.00	P	08/14/2025	18:05	LB136837
	Chromium	2.12	+/-5	U	10.0	P	08/14/2025	18:05	LB136837
	Copper	4.60	+/-10	U	20.0	P	08/14/2025	18:05	LB136837
	Lead	2.30	+/-6	U	12.0	P	08/14/2025	18:05	LB136837
	Nickel	3.06	+/-20	U	40.0	P	08/14/2025	18:05	LB136837
	Selenium	9.64	+/-10	U	20.0	P	08/14/2025	18:05	LB136837
	Silver	1.62	+/-5	U	10.0	P	08/14/2025	18:05	LB136837
	Zinc	16.7	+/-20	U	40.0	P	08/14/2025	18:05	LB136837