

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2732

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 09:35

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1242-1 (1)	5.82	5.81	5.71	5.91	-0.01
Aroclor-1242-2 (2)	5.84	5.83	5.73	5.93	-0.01
Aroclor-1242-3 (3)	5.90	5.89	5.79	5.99	-0.01
Aroclor-1242-4 (4)	6.00	5.99	5.89	6.09	-0.01
Aroclor-1242-5 (5)	6.73	6.72	6.62	6.82	-0.01
Tetrachloro-m-xylene	4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl	10.45	10.43	10.33	10.53	-0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2732

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 09:35

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1242-1 (1)	4.90	4.91	4.81	5.01	0.01
Aroclor-1242-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1242-3 (3)	5.08	5.08	4.98	5.18	0.01
Aroclor-1242-4 (4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1242-5 (5)	5.68	5.69	5.59	5.79	0.01
Tetrachloro-m-xylene	3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl	8.82	8.83	8.73	8.93	0.01