

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2732

Instrument ID: ECD_S

Calibration Date(s): 08/07/2025 08/07/2025

Calibration Times: 12:02 13:38

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031379.D</u>	CF 500 =	<u>PS031380.D</u>		
CF 750 =	<u>PS031381.D</u>	CF 1000 =	<u>PS031382.D</u>	CF 1500 =	<u>PS031383.D</u>		
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	18839000000	18369900000	17907800000	17569300000	16932500000	17923700000	4
2,4-D	3359960000	2835460000	2754100000	2719120000	2715430000	2876820000	10
2,4-DCAA	3597170000	3293120000	3103110000	3009470000	2867080000	3173990000	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2732

Instrument ID: ECD_S

Calibration Date(s): 08/07/2025 08/07/2025

Calibration Times: 12:02 13:38

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031379.D</u>	CF 500 =	<u>PS031380.D</u>		
CF 750 =	<u>PS031381.D</u>	CF 1000 =	<u>PS031382.D</u>	CF 1500 =	<u>PS031383.D</u>		
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	13728500000	13523000000	13089500000	12836400000	12402700000	13116000000	4
2,4-D	1540750000	1470040000	1417550000	1385470000	1322620000	1427290000	6
2,4-DCAA	909289000	862278000	831824000	818654000	804653000	845340000	5