

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2732
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031454.D **Time Analyzed:** 10:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.293	9.200	9.400	702.270	712.500	-1.4
2,4-D	8.417	8.324	8.524	683.000	705.000	-3.1
2,4-DCAA	7.290	7.195	7.395	706.660	750.000	-5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2732
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031454.D **Time Analyzed:** 10:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.897	9.803	10.003	643.260	712.500	-9.7
2,4-D	9.001	8.906	9.106	606.930	705.000	-13.9
2,4-DCAA	7.745	7.649	7.849	659.270	750.000	-12.1