

Surrogate Summary

SDG No.: Q2732

Client: ENTACT

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS031378.D	PIBLK-PS031378.D	2,4-DCAA	1	500	421	84		70 (61)	130 (136)
		2,4-DCAA	2	500	503	101		70 (61)	130 (136)
I.BLK-PS031453.D	PIBLK-PS031453.D	2,4-DCAA	1	500	506	101		70 (61)	130 (136)
		2,4-DCAA	2	500	540	108		70 (61)	130 (136)
PB169278BL	PB169278BL	2,4-DCAA	1	500	450	90		70 (61)	130 (136)
		2,4-DCAA	2	500	509	102		70 (61)	130 (136)
PB169278BS	PB169278BS	2,4-DCAA	1	500	561	112		70 (61)	130 (136)
		2,4-DCAA	2	500	509	102		70 (61)	130 (136)
PB169212TB	PB169212TB	2,4-DCAA	1	500	442	88		70 (61)	130 (136)
		2,4-DCAA	2	500	482	96		70 (61)	130 (136)
I.BLK-PS031460.D	PIBLK-PS031460.D	2,4-DCAA	1	500	516	103		70 (61)	130 (136)
		2,4-DCAA	2	500	554	111		70 (61)	130 (136)
Q2832-02MS	TG-S01MS	2,4-DCAA	1	500	774	155	*	70 (61)	130 (136)
		2,4-DCAA	2	500	876	175	*	70 (61)	130 (136)
Q2832-02MSD	TG-S01MSD	2,4-DCAA	1	500	780	156	*	70 (61)	130 (136)
		2,4-DCAA	2	500	880	176	*	70 (61)	130 (136)
I.BLK-PS031472.D	PIBLK-PS031472.D	2,4-DCAA	1	500	521	104		70 (61)	130 (136)
		2,4-DCAA	2	500	568	114		70 (61)	130 (136)
I.BLK-PS031475.D	PIBLK-PS031475.D	2,4-DCAA	1	500	487	97		70 (61)	130 (136)
		2,4-DCAA	2	500	504	101		70 (61)	130 (136)
Q2732-03	WC-A7-01-C	2,4-DCAA	1	500	584	117		70 (61)	130 (136)
		2,4-DCAA	2	500	655	131	*	70 (61)	130 (136)
I.BLK-PS031481.D	PIBLK-PS031481.D	2,4-DCAA	1	500	466	93		70 (61)	130 (136)
		2,4-DCAA	2	500	502	100		70 (61)	130 (136)