



Report of Analysis

Client: ENTACT Date Collected: 08/15/25

Project: 540 Degraw St, Brooklyn, NY - E9309 Date Received: 08/15/25

Client Sample ID: PIBLK-PD089925.D SDG No.: Q2732

Lab Sample ID: I.BLK-PD089925.D Matrix: TCLP

Analytical Method: 8081B % Solid: 0 Decanted:

Sample Wt/Vol: 1000 Units: mL Final Vol: 10000 uL

Soil Aliquot Vol: uL Test: TCLP Pesticide

Extraction Type: Injection Volume:

GPC Factor: 1.0 PH:

Prep Method: 3510C

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID

PD089925.D 1 08/15/25 pd081625

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|----------------------|-----------|-----------|---------------------|------------|---------|
| TARGETS | | | | | | |
| 58-89-9 | gamma-BHC (Lindane) | 0.0000037 | U | 0.0000037 | 0.000050 | mg/L |
| 76-44-8 | Heptachlor | 0.0000027 | U | 0.0000027 | 0.000050 | mg/L |
| 1024-57-3 | Heptachlor epoxide | 0.0000096 | U | 0.0000096 | 0.000050 | mg/L |
| 72-20-8 | Endrin | 0.0000032 | U | 0.0000032 | 0.000050 | mg/L |
| 72-43-5 | Methoxychlor | 0.000011 | U | 0.000011 | 0.000050 | mg/L |
| 8001-35-2 | Toxaphene | 0.00017 | U | 0.00017 | 0.0010 | mg/L |
| 57-74-9 | Chlordane | 0.000088 | U | 0.000088 | 0.00050 | mg/L |
| SURROGATES | | | | | | |
| 2051-24-3 | Decachlorobiphenyl | 23.9 | | 30 (57) - 150 (171) | 119% | SPK: 20 |
| 877-09-8 | Tetrachloro-m-xylene | 25.4 | | 30 (61) - 150 (148) | 127% | SPK: 20 |

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit