

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2732

Analytical Method: 8270E

Client: ENTACT

DataFile: BP025440.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169275BS	Pyridine	50	32.4	ug/L	65				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	37.8	ug/L	76				70 (76)	130 (103)	
	2-Methylphenol	50	39.0	ug/L	78				70 (69)	130 (109)	
	3+4-Methylphenols	50	37.6	ug/L	75				20 (67)	160 (106)	
	Hexachloroethane	50	38.7	ug/L	77				20 (76)	160 (118)	
	Nitrobenzene	50	40.1	ug/L	80				70 (58)	130 (106)	
	Hexachlorobutadiene	50	38.3	ug/L	77				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	41.0	ug/L	82				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	41.3	ug/L	83				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	40.3	ug/L	81				70 (60)	130 (115)	
	Hexachlorobenzene	50	39.8	ug/L	80				70 (73)	130 (106)	
	Pentachlorophenol	100	85.4	ug/L	85				20 (47)	160 (114)	