

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2732</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/13/2025 10:57</u>
Lab File ID:	<u>VN087525.D</u>	Init. Calib. Date(s):	<u>07/16/2025 07/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>17:05 18:54</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.717		7.98	20
1,1-Dichloroethene	0.571	0.567		-0.7	20
2-Butanone	0.615	0.665		8.13	20
Carbon Tetrachloride	0.502	0.519		3.39	20
Chloroform	1.251	1.367		9.27	20
Benzene	1.473	1.499		1.76	20
1,2-Dichloroethane	0.558	0.597		6.99	20
Trichloroethene	0.348	0.342		-1.72	20
Tetrachloroethene	0.322	0.306		-4.97	20
Chlorobenzene	1.123	1.142	0.3	1.69	20
1,2-Dichloroethane-d4	0.848	0.898		5.9	20
Dibromofluoromethane	0.345	0.337		-2.32	20
Toluene-d8	1.230	1.204		-2.11	20
4-Bromofluorobenzene	0.455	0.470		3.3	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.