

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN071625

Review By	Maresh Dadoda	Review On	7/17/2025 8:20:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:25:10 AM
SubDirectory	VN071625	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134794
Initial Calibration Stds	VP134795,VP134796,VP134797,VP134798,VP134799,VP134800
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134801
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087327.D	16 Jul 2025 16:10		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN087328.D	16 Jul 2025 17:05		JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN087329.D	16 Jul 2025 17:27	% d fail for com.#10 in 5 ppb	JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN087330.D	16 Jul 2025 17:49	LR- 10,20	JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN087331.D	16 Jul 2025 18:11	QR- 56	JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN087332.D	16 Jul 2025 18:32		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN087333.D	16 Jul 2025 18:54		JC\MD	Ok,M
8	IBLK	IBLK	VN087334.D	16 Jul 2025 19:16		JC\MD	Ok
9	VSTDICV050	ICVVN071625	VN087335.D	16 Jul 2025 19:59		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081325

Review By	John Carlone	Review On	8/14/2025 9:03:57 AM
Supervise By	Maresh Dadoda	Supervise On	8/18/2025 8:55:14 AM
SubDirectory	VN081325	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135108
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135109,VP135110

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087524.D	13 Aug 2025 09:04		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087525.D	13 Aug 2025 10:57	pH#Lot#V12668	JC\MD	Ok,M
3	VN0813MBL01	VN0813MBL01	VN087526.D	13 Aug 2025 11:19		JC\MD	Ok
4	VN0813WBL01	VN0813WBL01	VN087527.D	13 Aug 2025 11:41		JC\MD	Ok
5	VN0813WBS01	VN0813WBS01	VN087528.D	13 Aug 2025 12:39		JC\MD	Ok,M
6	PB169213TB	PB169213TB	VN087529.D	13 Aug 2025 13:01		JC\MD	Ok
7	VN0813WBSD01	VN0813WBSD01	VN087530.D	13 Aug 2025 13:23		JC\MD	Ok,M
8	Q2816-05DL	1057-MW-03A(17)DL	VN087531.D	13 Aug 2025 13:45	vial B pH<2	JC\MD	Ok
9	Q2816-01DL	1055-MW-01(23)DL	VN087532.D	13 Aug 2025 14:07	vial B pH<2	JC\MD	Ok
10	Q2816-08	1060-FB080725	VN087533.D	13 Aug 2025 14:29	vial B pH<2 FB	JC\MD	Ok
11	Q2825-02RE	FBRE	VN087534.D	13 Aug 2025 14:51	FB;For Confirmation	JC\MD	Not Ok
12	Q2825-01	TW1	VN087535.D	13 Aug 2025 15:14	vial B pH<2	JC\MD	Ok,M
13	Q2808-04RE	TP-7RE	VN087536.D	13 Aug 2025 15:36	vial B pH#5.0 Surrogate fail	JC\MD	Confirms
14	Q2827-04RE	TP-8RE	VN087537.D	13 Aug 2025 15:58	vial B pH#5.0 Surrogate fail	JC\MD	Confirms
15	Q2732-01	WC-A7-01-G	VN087538.D	13 Aug 2025 16:21	vial A pH#5.0	JC\MD	Ok
16	Q2832-02	TG-S01	VN087539.D	13 Aug 2025 16:43	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
17	Q2832-04	TG-S02	VN087540.D	13 Aug 2025 17:05	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
18	Q2832-06	TG-S03	VN087541.D	13 Aug 2025 17:27	vial A pH#5.0	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081325

Review By	John Carlone	Review On	8/14/2025 9:03:57 AM
Supervise By	Maresh Dadoda	Supervise On	8/18/2025 8:55:14 AM
SubDirectory	VN081325	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135108		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135109,VP135110		

19	Q2832-08	TG-S04	VN087542.D	13 Aug 2025 17:50	vial A pH#5.0	JC\MD	Ok
20	Q2832-10	TG-S05	VN087543.D	13 Aug 2025 18:12	vial A pH#5.0	JC\MD	Ok
21	Q2836-01	WC-A2-15-G	VN087544.D	13 Aug 2025 18:34	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
22	Q2836-05	WC-A2-16-G	VN087545.D	13 Aug 2025 18:56	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
23	Q2836-09	WC-A2-17-G	VN087546.D	13 Aug 2025 19:18	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
24	Q2836-13	WC-A5-02-G	VN087547.D	13 Aug 2025 19:40	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
25	Q2838-04	TP-11	VN087548.D	13 Aug 2025 20:02	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
26	Q2838-08	TP-10	VN087549.D	13 Aug 2025 20:24	vial A pH#5.0 Surrogate fail	JC\MD	ReRun
27	VSTDCCC050	VSTDCCC050EC	VN087550.D	13 Aug 2025 20:46		JC\MD	Ok,M

M : Manual Integration