

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2732</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ENTACT</u>	Datafile :	<u>VN087530.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0813WBSD01	Vinyl chloride	20	19.5	ug/L	98	6		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	18.9	ug/L	95	1		70 (74)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	0		40 (65)	160 (122)	20 (26)
	Carbon Tetrachloride	20	18.2	ug/L	91	2		70 (77)	130 (113)	20 (15)
	Chloroform	20	20.9	ug/L	104	5		70 (79)	130 (113)	20 (20)
	Benzene	20	18.9	ug/L	95	1		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	20.6	ug/L	103	1		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	17.3	ug/L	86	2		70 (77)	130 (113)	20 (15)
	Tetrachloroethene	20	16.6	ug/L	83	2		70 (67)	130 (123)	20 (15)
	Chlorobenzene	20	17.8	ug/L	89	5		70 (82)	130 (109)	20 (15)