

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2732

Lab File ID: VN087524.D

BFB Injection Date: 08/13/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:04

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.3
75	30.0 - 60.0% of mass 95	57
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.3
173	Less than 2.0% of mass 174	1.1 (1.6) 1
174	50.0 - 100.0% of mass 95	66
175	5.0 - 9.0% of mass 174	5.3 (8) 1
176	95.0 - 101.0% of mass 174	63.9 (96.8) 1
177	5.0 - 9.0% of mass 176	3.8 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087525.D	08/13/2025	10:57
VN0813WBL01	VN0813WBL01	VN087527.D	08/13/2025	11:41
VN0813WBS01	VN0813WBS01	VN087528.D	08/13/2025	12:39
VN0813WBSD01	VN0813WBSD01	VN087530.D	08/13/2025	13:23
WC-A7-01-G	Q2732-01	VN087538.D	08/13/2025	16:21