

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2733
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2719-01
Client ID:	50738DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		07/31/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2733
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2725-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	744		750		1	0.8		07/31/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: 001 willets Pt Blvd(july)DUP	SDG No.: Q2733 Sample ID: Q2730-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	205		196		1	4.79		07/30/2025